

# **ECAT FOUNDATION**

**External quality Control for Assays and Tests**

**With a focus on Thrombosis and Haemostasis**

## **REPORT**




### **SURVEY 2025-MIX2**

### **APTT/PT Mixing Test report**

### **Labcode 606**

Copyright © 2025  
ECAT Foundation

Date of Issue : 18-July-2025  
Survey : 2025-MIX2  
Report : APTT/PT Mixing Test

**Note:**

In the Survey Manual 2025 detailed information is given regarding the ECAT external quality assessment programme, including the statistical evaluation and explanation of the report.  
This Survey Manual 2025 should be considered as an integral part of this survey report.  
Please notice the information regarding the homogeneity of samples used and the between-laboratory variation in the paragraph on the statistical evaluation of the Survey Manual.

General InformationExclusion of results

Results < [value] or > [value] are excluded from the statistical analysis. When other results (e.g. deviating results) are excluded from the statistical analysis, these results are placed between brackets.

The following participant reported a deviating result which was excluded from the statistical analysis:

**Labcode 115:** For both samples a result of 0 is reported for the PT without incubation.

Participant "**labcode 1231**" reported deviating MRI values for both APTT and PT probably by mixing up the values. These MRI's were excluded and as a result of this no ratios were calculated.

Mean Reference Interval

For the interpretation of the mixing test results, the ratio of the mixing test over the mean of the reference interval (MRI) is calculated.

If a mixing test result for APTT or PT is reported, it is mandatory to report also the corresponding MRI. The MRI is the average value of the lower and upper limit of the reference interval.

Complaints

Any complaints regarding this survey report should be reported to the ECAT before **September, 23rd, 2025**. Complaints received after this date will not be taken into consideration.

This report is authorized by:

Dr P. Meijer  
Director

**Note:** The current Survey Manual can be downloaded from the 'Participant Area' of the ECAT website. Login and select the option 'View Documents'.

ECAT Foundation

Director: Dr. P. Meijer  
ECAT Office  
P.O. Box 107  
2250 AC Voorschoten, The Netherlands  
phone +31 (0) 71 3030 910; fax + 31 (0) 71 3030 919  
E-mail: info@ecat.nl  
Website: www.ecat.nl  
VAT number: NL802836872B01

Registration number with the Chamber of Commerce (KvK) Gouda : 41174102  
General terms of delivery are applicable to all our services.

All rights reserved. No part of this report may be reproduced, stored in a retrieval system, or by any means, electronic, mechanical, photocopying, recording or otherwise, without prior permission from the ECAT Foundation.  
Appendices are an integral part of the total report.