

ECAT FOUNDATION

External quality Control for Assays and Tests

With a focus on Thrombosis and Haemostasis

REPORT



SURVEY 2026-L1 Lupus Anticoagulant Labcode 606

Copyright © 2026
ECAT Foundation

Date of Issue : 10-May-2026

Survey : 2026-L1

Report : Lupus Anticoagulant

Note:

In the Survey Manual 2026 detailed information is given regarding the ECAT external quality assessment programme, including the statistical evaluation and explanation of the report.

This Survey Manual 2026 should be considered as an integral part of this survey report.

Please notice the information regarding the homogeneity of samples used and the between-laboratory variation in the paragraph on the statistical evaluation of the Survey Manual.

General Information**Exclusion of results**

Results < [value] or > [value] are excluded from the statistical analysis. When other results (e.g., deviating results) are excluded from the statistical analysis, these results are placed between brackets.

Participant (**labcode 1794**) has submitted three times the same results, suggesting that these are separate results. This is not the case and may affect the statistical analysis and therefore the second and third results were excluded.

Lupus Anticoagulant

When selecting the unit seconds; all results should be reported in seconds and not partly in ratios; e.g., the result for the ECAT sample, the result for normal plasma and the result for MRI.

Antiphospholipid Antibodies

Please be aware of the selection of the correct unit for the method group "Werfen Acustar / INOVA Quanta Flash". Since there is a difference in the order of magnitude between the results of the "Werfen Acustar / INOVA Quanta Flash" method group and the other methods, it is expressed in the report as CU/mL instead of U/mL.

Ratio Screen/Confirm

Several participants reported a value for the ratio screen/confirm which actually was not likely to be a ratio screen over confirm. Therefore these results were excluded from the statistical analysis.

Complaints

Any complaints regarding this survey report should be reported to the ECAT before **June 23rd 2026**. Complaints received after this date will not be taken into consideration.

This report is authorized by:

Dr. M.J. van Essen-Hollestelle
Programme Expert

Note: The current Survey Manual can be downloaded from the 'Participant Area' of the ECAT website. Login and select the option 'View Documents'.

ECAT Foundation

Director: Dr. P. Meijer
ECAT Office
P.O. Box 107
2250 AC Voorschoten, The Netherlands
phone +31 (0) 71 3030 910; fax + 31 (0) 71 3030 919
E-mail: info@ecat.nl
Website: www.ecat.nl
VAT number: NL802836872B01

Registration number with the Chamber of Commerce (KvK) Gouda : 41174102
General terms of delivery are applicable to all our services.

All rights reserved. No part of this report may be reproduced, stored in a retrieval system, or by any means, electronic, mechanical, photocopying, recording or otherwise, without prior permission from the ECAT Foundation.
Appendices are an integral part of the total report.